

Literature APPLICATION Guide

Medical Piezo Transducers (No Circuitry) IEC 60601-1-8 AMD2 (2020)



- Approved to IEC 60601-1-8 AMD2 (2020)
- UL File # MDAF2.E480572
- cUL File # MDAF8.E480572
- The user needs to apply a square wave signal at the rated drive frequency (200 Hz).
- This results in a sound with one frequency component < 1 kHz and at least 4 frequency components between 150 Hz and 4 kHz. Also, the relative sound pressure levels of the four frequency components with the largest sound pressure levels will be within 15 dB of each other.
- It is up to the user to adjust the square wave signal to meet the rise & fall times and pulse timings listed in Tables 3 & 4.

IEC60601-1-8 Piezo Transducer (No Circuitry)

Part Number	Typical Sound Level @ 10 cm	Freq (Hz)	Operating Voltage (Vp-p)	Current (mA)	Capacitance (pF)	Dia x Hgt (mm)	Housing Type
SBT200PC	90 dBa	200	1 - 30	1.5 - 10	70,000	44.5 x 14	2 PC Pins
SBT200FL	90 dBa	200	1 - 30	1.5 - 10	70,000	64.5 x 14	Flange w/Bare Wires

Call 317-612-1000 or Visit www.mspindy for More Details